

RAPICIDE™ PA

Spill Cleanup & Disposal | General Information

RAPICIDE™ PA Disinfectant is a single use high-level disinfectant for use in the MEDIVATORS ADVANTAGE PLUS™ Automated Endoscope Reprocessor or DSD EDGE™ Automated Endoscope Reprocessor. For more information on RAPICIDE PA Disinfectant please refer to the following documents:

RAPICIDE PA Disinfectant SDS forms for Part A and Part B:

- RAPICIDE PA Disinfectant Directions for Use
- RAPICIDE PA High-Level Disinfectant Safety, Efficacy and Microbiological Summary
- ADVANTAGE PLUS Automated Endoscope Reprocessor or DSD EDGE Automated Endoscope Reprocessor User Manual

ALWAYS read the RAPICIDE PA Disinfectant Part A and Part B labels before use along with the Directions for Use provided with the product.

ALWAYS inspect the box and containers for leaks or damages. Do NOT use leaking or damaged containers of RAPICIDE PA Disinfectant. If leaks or damage to the containers are found, please call Medivators Support at 1-800-444-4729 to report the incident.

While accidental spillage of RAPICIDE PA Disinfectant is rare, it may occur. It is important to follow all of the steps to ensure proper handling of the spill.

Spills of RAPICIDE PA Disinfectant

- 1. If a spill occurs when using a reprocessor, first CANCEL out of any cycle or program that is currently running.
- 2. When cancellation is complete, UNPLUG the reprocessor from the wall.
- 3. Put on PERSONAL PROTECTIVE EQUIPMENT (PPE) to include properly fitted dust- or splash-proof chemical safety goggles, face shield, protective gloves, protective apron, and rubber boots.
 - * Wear PPE for the entire spill cleanup procedure.
- 4. CONTAIN the spill.

If Part A was spilled:

- a) Slowly sprinkle sodium bicarbonate (baking soda) on the spill to neutralize, there will be a reaction (bubbling, etc.).
- A plastic stick or similar device can be used to help stir in the sodium bicarbonate.
- Add enough sodium bicarbonate so that the spill becomes the same thickness as pancake batter.
- d) Continue to sprinkle sodium bicarbonate until no more reaction occurs (bubbling stops). Once there is no longer a reaction, the spill has been neutralized.



NOTE: As a guide, for every 1 gallon of solution you spill, use 1/2 pound of sodium bicarbonate.

- e) Scoop up material and place in a disposal container.
 Absorb spillage to prevent material damage. Provide ventilation.
 Do not reuse the liquid material.
- f) Materials should be disposed of according to applicable local, state, and federal regulations.
- g) Run cold water for 15 minutes down any drains used.
- h) Thoroughly rinse off any equipment used with cold water.

If Part B was spilled:

For containment:

- a) Contain and/or absorb spill with inert material (e.g. sand, vermiculite), then place in a suitable container.
- b) Do not flush to sewer or allow to enter waterways.
- c) Use appropriate Personal Protective Equipment (PPE).

Methods for cleaning up:

a) Scoop up material and place in a disposal container.
 Absorb spillage to prevent material damage.

If Part A and Part B mixed was spilled:

- a) Flush the spilled solution to the drain using copious amounts of cold water if allowed by federal, state and/or the local regulatory authorities. If a drain is not immediately available, collect the spill by using mops, towels or similar material to clean up the spill.
- Thoroughly rinse and wring out any mops, towels or similar materials used. Materials should be disposed of according to applicable local, state, and federal regulations.
- c) Run cold water for 15 minutes down any drains used.
- d) Thoroughly rinse off any equipment used with cold water.
- Once spill is cleaned up and disposed of properly, plug in the reprocessor if necessary and follow instructions to reset up any cycles that were interrupted.





Leaking Bottles of RAPICIDE™ PA High-Level Disinfectant

ALWAYS inspect the box and containers for leaks or damages. Do NOT use leaking or damaged containers of RAPICIDE PA Disinfectant. If leaks or damage to the containers are found, please call Medivators Support at 1-800-444-4729 to report the incident. Follow the proper disposal instructions as indicated on the Safety Data Sheet and below to dispose of the leaking or damaged containers.

- 1. Put on PERSONAL PROTECTIVE EQUIPMENT (PPE) to include ANSI approved safety glasses or goggles, face shield, protective gloves, protective apron, and rubber boots.
 - * Wear PPE for the entire spill cleanup procedure.
- 2. Remove the leaking bottles from their current location and put them in a leak proof container such as a sink (with drain closed) or a mop bucket.
- 3. Follow the steps above under "CONTAIN the spill" to properly dispose of the leaking Part A or Part B solution.
- 4. If box is wet from a leaking container, thoroughly soak the box with water.
- 5. Never try to use, reuse, or save liquid material from a leaking container.



 \bigoplus

RAPICIDE™, ADVANTAGE PLUS™, and DSD EDGE™ are trademarks of Medivators Inc.



Manufactured in the USA by:





1A International Business Park #05-01 Singapore 609933 Tel: +65.6227.9698

Cantel Medical Asia/Pacific Pte. Ltd. Cantel Medical Devices (China) Co. Ltd. Unit 804-805, Innov Tower Block A, Hongmei Road, Xuhui 200233 Shanghai Tel: +86 21 60161380 Fax: +86 21 61210913

© 2018 Medivators Inc. P/N 50097-408/F

www.medivators.com

50097-408 REV E.indd 2 12/17/2018 9:26:27 AM

